

## AS/NZS 4760:2019 - Certificate of Compliance

**DrugCheckExpress SalivaTracer Multi-Drug Rapid Test Cassette (Oral Fluid)**  
**Manufactured for Royal Medical Supplies Pty Ltd.**  
**Cat# DCOFT-M-07-02 Lot #DOA2502018-S Expiry 2027-02-12**

The DrugCheckExpress SalivaTracer Multi-Drug Rapid Test Cassette (Oral Fluid) device Cat#: DCOFT-M-07-02 was tested in accordance with Appendix C3 – Verification of performance of devices used for the collection, on-site testing, transport and storage of oral fluid specimens – On-site testing devices, of the Australian and New Zealand Standard AS/NZS 4760:2019.

The DrugCheckExpress SalivaTracer Multi-Drug Rapid Test Cassette (Oral Fluid) device Cat#: DCOFT-M-07-02 was shown to satisfy the specified performance criteria when results are interpreted as described in the package insert (Number: RP5802300, Effective date: 2025-03-27).

Drug groups verified are amphetamine, methamphetamine, cocaine, opiates, oxycodone and Δ9-THC, using matrix matched oral fluid.

Benzodiazepines are verified at the cut-off level on the device (10ng/mL).

Results of the verification procedure used to assess compliance of the DrugCheckExpress SalivaTracer Multi-Drug Rapid Test Cassette (Oral Fluid) device Cat#: DCOFT-M-07-02 have been documented and are available to any requesting authority or accrediting agency.



David Batty  
Laboratory Director

31 March 2025

Accredited for compliance with ISO/IEC 17025 – Testing.

NATA is a signatory to the ILAC Mutual Recognition Arrangement for the mutual recognition of the equivalence of testing, medical testing, calibration, inspection, proficiency testing scheme providers and reference materials producers reports and certificates.

This certificate may not be reproduced except in full.

RASL Certificate of Compliance\_4, issue date 11 March 2025